Physikalisch-Technische Bundesanstalt



Braunschweig und Berlin

Member State of OIML Germany





OIML Certificate No. R60/2000-DE1-10.11

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Address: Person responsible: Physikalisch-Technische Bundesanstalt Bundesallee 100, 38116 Braunschweig Dr. Dirk Ratschko

Applicant

Name: Address Zemic Europe B.V. Leerlooierstraat 8

4871 EN Etten-Leur

Niederlande

Manufacturer of the certified type is the applicant.

Identification of the cer-Stra tified type

Strain gauge single point load cell

Type: L6T

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60, edition 2000 for accuracy classes C3, C4

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above. Physikalisch-Technische Bundesanstalt



OIML Certificate No. R60/2000-DE1-10.11

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Report and Test Reports

Test-Report No. 1.12-4044287 Test-Report No. 1.12-4044287-4 Report No. 1.12-4044287

The Issuing Authority

Dr. D. Ratschko Head of Department

18.10.2010



that includes 22 pages that includes 22 pages that includes 6 pages

The CIML Member

Dr. R. Schwartz Head of Division

18.10.2010

The load cells of the series L6T are single point load cells. They are made of aluminium and the strain gauge application is hermetically sealed.

The metrological characteristics for application in approved weighing instruments are listed in table 1.

Table 1: Essential data

Accuracy class			C	:3	C4
Maximum number of load cell intervals	nuc		3000		4000
Rated output		mV/V	2 .		
Maximum capacity	Enav	kg	50 / 75 / 100 / 150 / 200 / 250	300 / 500 / 635 / 1000	50 / 75 / 100 / 150 / 200 / 250
Minimum load cell verification interval	Vmin = (Emax / Y)		E _{max} / 6500	Emas / 15000	Emm. / 6500

Dead load: 0%-Emax. Safe overload: 150%-Emax. Input impedance: 409 Ω; Fraction: pLc = 0.7

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Reports is not permitted, although either may be reproduced in full.